



NITRR/S&P/MT/2020/585

Dated:01/10/2020

//Corrigendum//

With reference to Tender Enquiry no NITRR/S&P/MT/OT/2020/536 Dt: 10/09/2020 which is due on 07/10/2020, there is following amendment in tender:-

S. No	Point No in Technical Specification	Present Condition / clause/specification	Amended (read as)
1	20	All test blocks shall come with UKAS accredited calibration certificates	All Calibrated Hardness test blocks must be supplied along with calibration certificates according to ASTM E92 & ISO 6507-3 & 4545-3 standard.
2	20	The manufacturer should have its own hardness calibration laboratory certified by the national body for the accreditation of testing and calibration laboratories. All relevant details need to be submitted by the Supplier or Manufacturer. Copies of accreditation in accordance with recognized international standard ISO 17025 for competence of testing and calibration laboratories shall be provided.	The manufacturer should have its own hardness calibration laboratory certified by national body for the accreditation of testing and calibration laboratories. All relevant details need to be submitted by the Supplier or Manufacturer. Copies of accreditation in accordance with recognized international standard ISO 17025 or NABL India for the competence of testing and calibration laboratories shall be provided.

Kindly submit your bid as per above within the following revised schedule of tender:

Last Date and Time for submission of Bids :10/10/2020 Upto 18:00Hrs

Date and Time for Opening of Bids :12/10/2020 At 16:00 Hrs

All other terms & conditions remain unchanged.

-sd-
Registrar
NIT Raipur