



राष्ट्रीय प्रौद्योगिकी संस्थान रायपुर
NATIONAL INSTITUTE OF TECHNOLOGY RAIPUR
(An Institute of National Importance)
G.E. Road, Raipur – 492010 (CG)

Phone: (0771) 4270619, Fax : (0771) 4061205, Email: registrar@nitrr.ac.in ,Website: www.nitrr.ac.in

No.....92..... /NITRR/TEQIP-III/2018

Raipur, Date 12/02/2018

Corrigendum

Corrigendum in the Technical Specification, Warranty Clause and Last date of Submission of Tender for Equipment– Gas Chromatography Mass Spectrometry (GCMS) under TEQIP-III (BID REFERENCE NO: TEQIP-III/nitr/26)

Following corrections are to be considered in **Technical specification, Warranty Clause and Last date of submission** of the above Tender:

Technical specification :

1. Existing

Column oven working temperature:	Room temperature + 4°C-450°C or better
----------------------------------	--

Change to

Column oven working temperature:	+ 4°C to 450°C or better
----------------------------------	--------------------------

2 .Existing

Detector:	High energy dynode electron multiplier Detectors such as Flame Ionization Detector, Thermal Conductivity Detector, Electron Capture Detector, Nitrogen Phosphorus Detector, Flame Photometric Detector full range of Thermo Scientific mass spectrometers • Maximum number Installed: 3 including a mass spectrometers on the same oven
-----------	--

Change to

Detector:	High energy electron multiplier Detectors such as Flame Ionization Detector, Thermal Conductivity Detector, Electron Capture Detector, Nitrogen Phosphorus Detector, Flame Photometric Detector etc. • Minimum number Installed: 3 including a mass spectrometers on the same oven
-----------	---